



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

1 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
Permanent Facility					
1	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @ 10 kHz	By using Reference Multimeter Fluke-8508A by Direct Method	1 A to 20 A	0.11 % to 0.30 %
2	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @ 50 Hz to 1 kHz,	By using Reference Multimeter Fluke-8508A by Direct Method	1 A to 20 A	0.10 % to 0.12 %
3	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @ 50 Hz to 10 kHz,	By using Reference Multimeter Fluke-8508A by Direct Method	100 µA to 1 mA	0.08 % to 0.058 %
4	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @ 50 Hz to 10 kHz,	By using Reference Multimeter Fluke-8508A by Direct Method	100 mA to 1 A	0.060 % to 0.11 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

2 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
5	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @ 50 Hz to 10 kHz,	By using Reference Multimeter Fluke-8508A by Direct Method	1 mA to 100 mA	0.058 % to 0.060 %
6	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @10 Hz,	By using Reference Multimeter Fluke-8508A by Direct Method	100 mA to 1 A	0.06 % to 0.11 %
7	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @10 Hz,	By using Reference Multimeter Fluke-8508A by Direct Method	100 $\mu$ A to 100 mA	0.08 % to 0.06 %
8	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Power & Active Energy Single Phase & Three Phase @ 50 Hz, 240 V/ 1 mA to 100 A at UPF, 0.5PF & 0.8 PF	By using Reference Energy Meter ZERA-MT310 by Comparison Method	0.12 W to 72 kW	0.076%
9	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Resistance @ 1 kHz	By using RLC Digibridge GenRad 1658 by Direct Method	100 Ohm to 10 k Ohm	0.12 % to 0.42 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

3 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
10	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Resistance @1 kHz	By using RLC Digibridge GenRad 1658 by Direct Method	1 Ohm to 100 Ohm	0.12%
11	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 10 Hz	By using Reference Multimeter Fluke-8508A by Direct Method	10 mV to 10 V	0.08 % to 0.036 %
12	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 10 Hz	By using Reference Multimeter Fluke-8508A by Direct Method	10 V to 100 V	0.036 % to 0.044 %
13	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 10 Hz,	By using Reference Multimeter Fluke-8508A by Direct Method	1 mV to 10 mV	0.93 % to 0.08 %
14	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 10 kHz	By using Reference Multimeter Fluke-8508A by Direct Method	1 mV to 10 mV	0.62 % to 0.069 %





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

4 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
15	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 10 kHz,	By using Reference Multimeter Fluke-8508A by Direct Method	10 mV to 100 mV	0.069 % to 0.017 %
16	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 10 kHz,	By using Reference Multimeter Fluke-8508A by Direct Method	100 mV to 100 V	0.017 % to 0.013 %
17	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 100 kHz,	By using Reference Multimeter Fluke-8508A by Direct Method	10 mV to 100 mV	0.33 % to 0.12 %
18	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 100 kHz,	By using Reference Multimeter Fluke-8508A by Direct Method	100 mV to 10 V	0.12 % to 0.031 %
19	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 50 Hz to 1 kHz,	By using Reference Multimeter Fluke-8508A by Direct Method	1 mV to 100 mV	0.57 % to 0.02 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

5 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
20	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 50 Hz to 1 kHz,	By using Reference Multimeter Fluke-8508A by Direct Method	100 mV to 100 V	0.020 % to 0.013 %
21	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 50 Hz to 10 kHz	By using Reference Multimeter Fluke-8508A by Direct Method	100 V to 1000 V	0.013 % to 0.017 %
22	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	Capacitance @ 1 kHz	By using Decade Capacitance Box Vaiseshikha-Type 7500 by Direct Method	10 nF	1.16 %
23	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	Capacitance@1 kHz	By using RLC Digibridge GenRad 1658 by Direct/ Comparison Method	1 nF to 100 $\mu$ F	0.12 %
24	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	Inductance @1 kHz	By using RLC Digibridge GenRad 1658 by Direct / Comparison Method	1 mH to 10 H	0.19 % to 0.13 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

6 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)( $\pm$ )
25	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	Inductance @1 kHz	By Using RLC Digibridge GenRad 1658 by Direct Method	100 $\mu$ H to 1 mH	0.32 % to 0.19 %
26	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Current @ 50 Hz to 1 kHz	Using Multifunction Calibrator,Fluke-550 2 A By Direct Method	30 $\mu$ A to 100 $\mu$ A	0.56 % to 0.29 %
27	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Current @ 50 Hz to 1 kHz	By using MFC 5502A by Direct / comparison Method	1 A to 10 A	0.08 % to 0.14 %
28	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Current @ 50 Hz to 1 kHz	By using MFC 5502A by Direct/ comparison method	10 A to 20 A	0.14 % to 0.2 %
29	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Current @ 50 Hz to 1 kHz	By using MFC 5502A by Direct / comparison method	100 mA to 1 A	0.075 % to 0.05 %





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

7 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
30	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Current @ 50 Hz to 1 kHz	By using MFC 5502A Direct method	100 mA to 1 A	0.075 % to 0.08 %
31	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Current @ 50 Hz,	Using Multifunction Calibrator,Fluke-550 2 A & Current Coil, Fluke By Direct Method.	20 A to 1000 A	0.82 % to 0.85 %
32	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Current @ 50Hz to 1 kHz,	Using Multifunction Calibrator,Fluke-550 2 A By Direct Method.	1 mA to 100 mA	0.14 % to 0.08 %
33	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Current @ 50Hz to 1kHz	Using Multifunction Calibrator,Fluke-550 2 A By Direct Method.	100 µA to 1 mA	0.29 % to 0.14 %
34	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Power (1 Phase) @UPF, 50 Hz, 30V to 320V/ 1 mA to 20A	Using Multifunction Multiproduct Calibrator ,Fluke-5502 A +50 Turn By Direct Method.	30 mW to 6.4 kW	0.24 % to 0.07 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

8 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
35	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Voltage @ 50Hz to 1kHz,	Using Multifunction Calibrator,Fluke-550 2 A By Direct Method.	1 mV to 10 mV	2.5 % to 0.36 %
36	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Voltage @ 50Hz to 1kHz,	Using Multifunction Calibrator, Fluke -5502A By Direct Method.	10 mV to 100 mV	0.36 % to 0.09 %
37	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Voltage @ 50Hz to 1kHz,	Using Multifunction Calibrator ,Fluke -5502A By Direct / Comparison Method.	10 V to 1000 V	0.060 % to 0.064 %
38	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Voltage @ 50Hz to 1kHz,	Using Multifunction Calibrator ,Fluke -5502A By Direct Method.	100 mV to 10 V	0.09 % to 0.06 %
39	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	Capacitance @ 1 kHz	By using Decade Capacitance Box Vaiseshikha-Type 7500 by Direct Method	1 nF	1.16 %
40	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	Capacitance @ 1 kHz	By using Decade Capacitance Box Vaiseshikha-Type 7500 by Direct Method	100 nF	1.16 %





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

9 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
41	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	Capacitance @ 1 kHz,	By using Decade Capacitance Box Vaiseshikha-Type 7500 by Direct / Comparison Method	1 $\mu$ F	1.16 %
42	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	Capacitance @ 100 Hz	By using Decade Capacitance Box Vaiseshikha-Type 7500 by Direct Method	1 $\mu$ F	1.16 %
43	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	Capacitance @ 100 Hz	By using Decade Capacitance Box Vaiseshikha-Type 7500 by Direct Method	1 mF	1.17 %
44	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	Capacitance @ 100 Hz	By using Decade Capacitance Box Vaiseshikha-Type 7500 by Direct Method	10 $\mu$ F	1.16 %
45	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	Capacitance @ 100 Hz	By using Decade Capacitance Box Vaiseshikha-Type 7500 by Direct Method	100 $\mu$ F	1.17%
46	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	Capacitance@ 1kHz	By using MFC Fluke 5502A by Direct Method	1 nF to 100 $\mu$ F	2.0 % to 1.91 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

10 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
47	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	Inductance @1 kHz	using Decade inductance box Vaiseshika/7500 by Direct method	1 H	1.16 %
48	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	Inductance @1 kHz	By using Decade inductance box Vaiseshika/7500 by Direct method	1 mH	1.16 %
49	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	Inductance @1 kHz	using Decade inductance box Vaiseshika/7500 by Direct/ Comparison method	10 H	1.16 %
50	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	Inductance @1 kHz	using Decade inductance box Vaiseshika/7500 by Direct/ Comparison method	10 mH	1.16 %
51	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	Inductance @1 kHz	By using Decade inductance Box Vaiseshikha/ 7500 by Direct Method	100 µH	1.16 %
52	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	Inductance @1 kHz	using Decade inductance box Vaiseshika/7500 by Direct method	100 mH	1.16 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

11 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
53	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	Power Factor	Using Multifunction MultiProduct Calibrator,Fluke-550 2A By Direct Method	50Hz, 0.1 pF Lead-UPF to 0.1 pF Lag	0.02pF
54	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	AC Capacitance @1kHz	Using Multifunction Reference Multimeter,Fluke-88 46A By Direct Method	1 nF to 10 nF	5.2 % to 1.75 %
55	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	AC Capacitance @1kHz	Using Multifunction Reference Multimeter,Fluke-88 46A By Direct Method.	10 nF to 10 mF	1.75 % to 2.21 %
56	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	DC Current	By using Reference Multimeter Fluke-8508A by Direct Method	1 A to 10 A	0.023 % to 0.09 %
57	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	DC Current	By using Reference Multimeter Fluke-8508A by Direct Method	10 µA to 10 mA	0.009 % to 0.0029 %
58	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	DC Current	By using Reference Multimeter Fluke-8508A by Direct Method	10 A to 20 A	0.051 % to 0.049 %





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

12 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)( $\pm$ )
59	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Current	By using Reference Multimeter Fluke-8508A by Direct Method	10 mA to 100 mA	0.0029 % to 0.007 %
60	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Current	By using Reference Multimeter Fluke-8508A by Direct Method	100 mA to 1 A	0.007 % to 0.023 %
61	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	Using Reference Multimeter Fluke 8508A By Direct Method	0.1 mV to 1 mV	0.13 % to 0.016 %
62	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	By using Reference Multimeter Fluke 8508A by Direct Method	1 mV to 10 mV	0.016 % to 0.004 %
63	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	By using Reference Multimeter Fluke 8508 A by Direct Method	1 V to 1000 V	0.0007 % to 0.001 %
64	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	using Reference Multimeter Fluke -8508A By Direct Method	10 mV to 100 mV	0.004 % to 0.001 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

13 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)( $\pm$ )
65	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	By using Reference Multimeter Fluke-8508A by Direct Method	100 mV to 1 V	0.001 % to 0.0007 %
66	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance	By using Reference Multimeter Fluke-8508A by Direct Method	1 M ohm to 10 M ohm	0.002 % to 0.0043 %
67	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance	By using Reference Multimeter Fluke-8508A by Direct / Comparison Method	1 m ohm to 10 m ohm	0.47 % to 0.049 %
68	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance	By using Reference Multimeter Fluke-8508A by Direct Method	1 ohm to 10 ohm	0.0063 % to 0.0017 %
69	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance	By using Reference Multimeter Fluke-8508A by Direct Method	10 k ohm to 1 M ohm	0.001 % to 0.002 %
70	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance	By using Reference Multimeter Fluke-8508A by Direct Method	10 m ohm to 100 m ohm	0.052 % to 0.0071 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

14 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
71	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance	By using Reference Multimeter Fluke-8508A by Direct Method	10 M ohm to 100 M ohm	0.0043 % to 0.026 %
72	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance	By using Reference multimeter Fluke 8508A by Direct Method	10 Ohm to 10 k Ohm	0.0017 % to 0.001 %
73	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance	By using Reference Multimeter Fluke-8508A by Direct Method	100 m ohm to 1 ohm	0.0071 % to 0.0063 %
74	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance	By using Reference Multimeter Fluke-8508A by Direct Method	100 M ohm to 10 G ohm	0.026 % to 0.29 %
75	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using Multifunction Calibrator Fluke-5502A By Direct/ Comparison Method	1 A to 10 A	0.050 % to 0.075 %
76	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using Multifunction Calibrator - Fluke-5502A By Direct Method.	1 mA to 100 mA	0.02 % to 0.017 %





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

15 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
77	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using Multifunction Calibrator,Fluke-550 2A By Direct Method	10 $\mu$ A to 100 $\mu$ A	0.26 % to 0.041 %
78	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using Multifunction Calibrator Fluke-5502A By Direct Method.	10 A to 20 A	0.089 % to 0.13 %
79	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using Multifunction Calibrator ,Fluke -5502A By Direct Method	100 $\mu$ A to 1 mA	0.041 % to 0.02 %
80	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using Multifunction Calibrator,Fluke-550 2A By Direct Method.	100 mA to 1 A	0.017 % to 0.069 %
81	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using MFC 5502A +50 Turn Current Coil by Direct Method	20 A to 1000 A	0.91 % to 0.85 %
82	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Resistance	By using MFC Fluke 5502A by Direct Method	1 M Ohm to 10 M Ohm	0.019 % to 0.07 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

16 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
83	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Resistance	Using Multifunction Calibrator ,Fluke -5502A By Direct Method.	1 ohm to 10 ohm	0.58 % to 0.069 %
84	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Resistance	Using Multifunction Calibrator ,Fluke -5502A By Direct / Comparison Method.	10 M ohm to 100 M ohm	0.070 % to 0.58 %
85	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Resistance	Using Multi function Calibrator,Fluke -5502A By Direct Method	10 ohm to 100 k ohm	0.06 % to 0.014 %
86	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Resistance	Using Multifunction Calibrator ,Fluke -5502A By Direct Method.	100 k ohm to 1 M ohm	0.014 % to 0.019 %
87	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Resistance	Using Multifunction Calibrator ,Fluke -5502A By Direct Method.	100 M ohm to 1 G ohm	0.58 % to 1.75 %
88	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Resistance	By using MFC Fluke 5502A by Direct Method	100 m Ohm to 1 Ohm	0.8 % to 0.58 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

17 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)( $\pm$ )
89	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Voltage	using Multifunction calibration Fluke 5502A direct Method	1 mV to 10 mV	0.36 % to 0.043 %
90	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Voltage	using Multifunction calibrator,Fluke-550 2A By Direct Method	1 V to 100 V	0.064 % to 0.0072 %
91	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Voltage	using Multifunction calibrator,Fluke-550 2A By Direct Method	10 mV to 100 mV	0.043 % to 0.011 %
92	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Voltage	using Multifunction calibrator,Fluke-550 2A By Direct Method	100 mV to 1 V	0.011 % to 0.064 %
93	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Voltage	using Multifunction calibrator,Fluke-550 2A By Direct/Comparison Method	100 V to 1000 V	0.0072 % to 0.0066 %





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

18 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
94	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	High Resistance	By using Decade Resistor Box Vaiseshikha-Type 8400 HVSD(Standard) by Direct Method	1 G ohm	1.41 %
95	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	High Resistance	By using Decade Resistor Box Vaiseshikha-Type 8400 HVSD(Standard) by Direct Method	1 k ohm	0.6 %
96	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	High Resistance	By using Decade Resistor Box Vaiseshikha-Type 8400 HVSD(Standard) by Direct Method	1 M Ohm	1.0 %
97	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	High Resistance	By using Decade Resistor Box Vaiseshikha-Type 8400 HVSD(Standard) by Direct Method	1 T ohm	8.33 %
98	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	High Resistance	By using Decade Resistor Box Vaiseshikha-Type 8400 HVSD(Standard) by Direct Method	10 G ohm	1.7 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

19 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
99	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	High Resistance	By using Decade Resistor Box Vaiseshikha-Type 8400 HVSD(Standard) by Direct Method	10 k ohm	0.06 %
100	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	High Resistance	By using Decade Resistor Box Vaiseshikha-Type 8400 HVSD(Standard) by Direct Method	10 M OHM	1.4 %
101	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	High Resistance	By using Decade Resistor Box Vaiseshikha-Type 8400 HVSD(Standard) by Direct Method	100 G ohm	3.8 %
102	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	High Resistance	By using Decade Resistor Box Vaiseshikha-Type 8400 HVSD(Standard) by Direct Method	100 k ohm	0.6 %
103	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	High Resistance	By using Decade Resistor Box Vaiseshikha-Type 8400 HVSD(Standard) by Direct Method	100 M ohm	1.4 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

20 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
104	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Low Resistance	By using Micro/Milli Ohm Meter Calibrator Vaiseshikha-9409 CAL by Direct Method	1 m ohm	0.16 %
105	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Low Resistance	By using Micro/Milli Ohm Meter Calibrator Vaiseshikha-9409 CAL by Direct Method	1 ohm	0.07 %
106	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Low Resistance	By using Micro/Miliohm meter calibrator Vaiseshika-9409 Cal by Direct Method	10 $\mu$ Ohm	3.2 %
107	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Low Resistance	By using Micro/Milli Ohm Meter Calibrator Vaiseshikha-9409 CAL by Direct Method	10 $\mu$ ohm	3.48 %
108	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Low Resistance	By using Micro/Milli Ohm Meter Calibrator Vaiseshikha-9409 CAL by Direct Method	10 m ohm	0.08 %





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

21 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
109	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Low Resistance	By using Micro/Milli Ohm Meter Calibrator Vaiseshikha-9409 CAL by Direct/Comparison Method	100 $\mu$ ohm to	1.6 %
110	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Low Resistance	By using Micro/Milli Ohm Meter Calibrator Vaiseshikha-9409 CAL by Direct Method	100 m ohm	0.07 %
111	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Low Resistance	By using Micro/Milli Ohm Meter Calibrator Vaiseshikha-9409 CAL by Direct Method	2 ohm	0.07 %
112	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance	By using Decade Resistor Box Vaiseshikha-Type 8400 HVSD(Standard) by Direct Method	100 M Ohm	1.4 %
113	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	B -Type	Using Multiproduct Calibrator ,Fluke-5520A, Using ITS-90 Scale by Simulation Method	600 °C to 1800 °C	0.83 °C



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

22 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)( $\pm$ )
114	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	E -Type TC	Using Multiproduct Calibrator ,Fluke -5520A,Using ITS -90 ScaleV Method by Simulation Method	(- )100 °C to 1000 °C	0.66 °C
115	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	J -Type TC	Using Multiproduct Calibrator Fluke -5520A Using ITS-90 Scale by Simulation Method	(- )200 °C to 1200 °C	0.45 °C
116	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	K -Type TC	Using Multiproduct Calibrator Fluke -5520A Using ITS-90 Scale by Simulation Method	(- )200 °C to 1370 °C	0.61 °C
117	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	N -Type TC	Using Multiproduct Calibrator ,Fluke -5520 A ,Using ITS-90 Scale by Simulation Method	(- )200 °C to 1300 °C	0.47°C
118	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	R -Type TC	Using Multiproduct Calibrator ,Fluke -5520A,Using ITS-90 scale by Simulation Method	2 °C to 1769 °C	0.67°C



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

23 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)( $\pm$ )
119	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	RTD (PT-100)	Using Multiproduct calibrator,Fluke 5520A Using ITS-90 Scale for ohm &mV Method by Simulation Method	(-) 200 °C to 800 °C	0.39 °C
120	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	S -Type TC	Using Multiproduct Calibrator ,Fluke -5520A,Using ITS-90 scale by Simulation Method	2 °C to 1769 °C	0.84 °C
121	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	T -Type TC	Using Multiproduct Calibrator ,Fluke -5520 A ,Using ITS-90 Scale by Simulation Method	(-)200 °C to 400 °C	0.80 °C
122	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Digital/Analogue stop watch, Digital Timer	Using Time Interval Meter by Comparison Method	1 hr to 5 hr	6.65 s to 24.62 s
123	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Digital/Analogue stop watch, Digital Timer	Using Time Interval Meter by Comparison Method	1 s to 60 s	0.14 s to 0.16 s





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

24 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)( $\pm$ )
124	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Digital/Analogue Stop watch, Digital Timer	Using Time Interval Meter by Comparison Method	200 ms to 1 s	0.07 s to 0.14 s
125	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Digital/Analogue stop watch, Digital Timer	Using Time Interval Meter by Comparison Method	5 hr to 27 hr	24.62 s to 26 s
126	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Digital/Analogue stop watch, Digital Timer	Using Time Interval Meter by Comparison Method	60 s to 1 hr	0.16 s to 6.65 s
127	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Frequency measurement for MFC's, DMM	By using Reference Multimeter Fluke-8508A by Direct / Comparison Method	10 Hz to 1 MHz	0.0013 % to 0.0012 %
128	ELECTRO-TECHNICAL-TIME & FREQUENCY (Source)	Frequency meter, DMM	Using Multifunction Multiproduct Calibrator, Fluke 5502A By Direct Method.	10 Hz to 1 MHz	0.062 % to 0.0021 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

25 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
129	MECHANICAL-ACCELERATION AND SPEED	RPM Indicator	Using Digital Tachometer & Mechanical type Tachometer Calibrator By Comparison Method	10 rpm to 20000 rpm	2.0rpm
130	MECHANICAL-ACCELERATION AND SPEED	Tachometer (contact Mode)	Using Digital Tachometer & Mechanical type Tachometer Calibrator By Comparison Method	10 rpm to 1000 rpm	3.4rpm
131	MECHANICAL-ACCELERATION AND SPEED	Tachometer (contact Mode)	Using Digital Tachometer & Mechanical type Tachometer Calibrator By Comparison Method	1000 rpm to 5000 rpm	6.6rpm
132	MECHANICAL-ACCELERATION AND SPEED	Tachometer (Noncontact Type)	Using Digital Tachometer & Mechanical type Tachometer Calibrator By Comparison Method	10 rpm to 100 rpm	3rpm
133	MECHANICAL-ACCELERATION AND SPEED	Tachometer (Noncontact Type)	Using Digital Tachometer & Mechanical type Tachometer Calibrator By Comparison Method	100 rpm to 10000 rpm	14.6rpm



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

26 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)( $\pm$ )
134	MECHANICAL-ACCELERATION AND SPEED	Tachometer (Noncontact Type)	Using Digital Tachometer & Mechanical type Tachometer Calibrator By Comparison Method	10000 rpm to 90000 rpm	15.0rpm
135	MECHANICAL-ACOUSTICS	Sound level meter @ 1 kHz	By using Sound Level calibrator by Direct/Comparison Method	94 dB	0.58 dB
136	MECHANICAL-ACOUSTICS	Sound Level Meter @1kHz	By using Sound Level Calibrator by Direct/Comparison Method	114 dB	0.58 dB
137	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Bevel Protractor / Combination set/ Angle Protractor L.C. 1 Min	Using Angle Gauge Set	0 to 360 °	5.0min of arc.
138	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Bore Gauge (Transmission Movement Only) L.C 0.001 mm	Using Length Measuring Machine By Comparison Method.	0 to 1 mm	1 $\mu$ m





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

27 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
139	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	caliper (vernier/Dial/Digital) L.C 0.01mm	using caliper checker/Long Gauge Blocks By Comparison Method	Up to 600 mm	13µm
140	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	caliper (vernier/Dial/Digital) L.C 0.01mm	using Caliper Checker /Long Gauge Blocks By Comparison Method	up to 1000 mm	18µm
141	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Coating Thickness Gauge L.C. 0.001mm	Using Standard Plastics Foils By Comparison Method	Up to 732 µm	1.2µm
142	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Cylindrical Measuring Pins	Using Length Measuring Machine By comparison Method	0.1 mm to 20 mm	1.0µm
143	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Depth Caliper (Vernier/Dial/Digital) L.C 0.01mm	Using Gauge Block set,Long Gauge Block, & Surface Plate By Comparison Method	up to 300 mm	9µm



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

28 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
144	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Depth Micrometer (Traverse) L.C 0.01 mm	Using Gauge Block Set, Long Gauge Block Set & Surface Plate By Comparison Method as per JIS B 7544	up to 150 mm	8.5µm
145	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Dial Gauge (Lever Type) L.C 0.001mm	Using Length Measuring Machine By Comparison Method	0 to 0.14 mm	1µm
146	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Dial Gauge (Lever Type) L.C 0.01 mm	Using Length Measuring Machine By Comparison Method	0 to 1 mm	6µm
147	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Dial Gauge (Plunger Type) L.C 0.001mm	Using Length Measuring Machine By Comparison Method	0 to 25 mm	1µm
148	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Dial Thickness Gauge L.C 0.001 mm	Using Slip Gauge Set By Comparison Method.	0 to 50 mm	1µm



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

29 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
149	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	External Micrometer L.C 0.001 mm	Using Gauge Block Set,Long Gauge Block By Comparison Method	25 mm to 300 mm	4.7µm
150	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	External Micrometer L.C 0.001 mm	Using Gauge Block Set,Long Gauge Block By Comparison Method	up to 25 mm	1µm
151	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	External Micrometer L.C 0.01 mm	Using Gauge Block Set,Long Gauge Block By Comparison Method	300 mm to 500 mm	10µm
152	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Feeler Gauge	Using Length Measuring Machine By Comparison Method	0.03 mm to 1 mm	1µm
153	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Height Gauge (Vernier/Dial/Digital) L.C. 0.01mm	Using Caliper Checker,Surface Plate By Comparison Method	up to 600 mm	13µm





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

30 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
154	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Internal Micrometer (Stick Type) L.C 0.01 mm	Using LMM,Long Gauge Block,Comparator Dial By Comparison Method	Head Travel 50.0 mm to 63 mm	6µm
155	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Internal Micrometer (Stick Type) L.C 0.01 mm	Using LMM,Long Gauge Block,Comparator Dial By Comparison Method	Overall length 50 to 30 mm	11µm
156	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Measuring Scale L.C. 1 mm	Using Tape & Scale Measuring Machine By Comparison Method	up to 1000 mm	290vL µm (L in mtr)
157	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Measuring Tape L.C. 1 mm	Using Tape & Scale Measuring Machine By Comparison Method based on IS 1269	up to 50000 mm	290 v L ( L in mtr)µm
158	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Micrometer Setting Standard	Using Length Measuring Machine,Gauge Block Set,Comparator Dial By Comparison Method.	100 mm to 300 mm	4.9µm



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

31 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
159	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Micrometer Setting Standard	Using Length Measuring Machine,Gauge Block Set ,Comparator Dial By Comparison Method.	25 mm to 100 mm	2µm
160	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Pistol Caliper L.C 0.1 mm	Using Slip Gauge Set By Comparison Method	0 to 50 mm	70µm
161	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Plain Plug Gauge / OD Gauge	Using Length Measuring Machine By Comparison Method	100 mm to 300 mm	4.9µm
162	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Plain Plug Gauge / OD Gauge	Using Length Measuring Machine By Comparison Method	3 mm to 100 mm	2.0µm
163	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Plain Ring Gauge	Using Length Measuring Machine By Comparison Method	100 mm to 300 mm	4.4µm



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

32 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
164	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Plain Ring Gauge	Using Length Measuring Machine By Comparison Method	3 mm to 100 mm	2.0µm
165	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Snap Gauge / Gap Gauge	Using Length Measuring Machine By Comparison Method	3 mm to 50 mm	1.2µm
166	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Snap Gauge / Gap Gauge	Using Length Measuring Machine By Comparison Method	50 mm to 300 mm	4.0µm
167	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Thickness Gauge	Using Length Measuring Machine By Comparison Method.	0.01 mm to 5 mm	1µm
168	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Thread Plug Gauge (Effective Diameter)	Using Length Measuring Machine Setting Disc, Standard Wires By Comparison Method	3 mm to 100 mm	1.3µm





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

33 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
169	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Thread Plug Gauge (Effective Diameter)	Using Length Measuring Machine Setting Disc,Standard Wires By Comparison Method	100 mm to 300 mm	5.0µm
170	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Thread Ring Gauge (Effective Diameter)	Using Length Measuring Machine ,Master Setting Ring By Comparison Method	5 mm to 100 mm	3.20µm
171	MECHANICAL-PRESSURE INDICATING DEVICES	Digital Pressure Indicator ,Manometer Magnahelic Gauge ,Low Pressure Gauge ,Pressure Transmitter,Pressure Switches Differential Pressure Gauges /Indicators - (Pneumatic Pressure)	Using Digital Pressure indicator, Calibrator Make: Polltech Instrument & Test pump , 6.5 Digit DMM Calibrator by Comparison method DKD-R6-01	0 to 200 mbar	0.1mbar



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

34 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
172	MECHANICAL-PRESSURE INDICATING DEVICES	Digital/Analog Pressure Gauges,Indicators,Pressure Transmitter,Pressure Switch,Differential Pressure Gauge/Indicators - (Pneumatic Pressure)	Using Digital Pressure Gauge & Pneumatic Test Pump, 6.5 Digit DMM Calibrator by Comparison Method DKD-R6-01	0 to 2 bar	0.0006bar
173	MECHANICAL-PRESSURE INDICATING DEVICES	Digital/Analog Pressure Gauges,Pressure Indicators,Pressure Transmitter,Pressure Switches -(Hydraulic Pressure)	Using Digital Pressure Gauge & Pneumatic Test Pump , 6.5 Digit DMM Calibrator By Comparison Method DKD-R6-01	1 bar to 350 bar	0.07bar
174	MECHANICAL-PRESSURE INDICATING DEVICES	Digital Pressure Indicator ,Manometer Magnahelic Gauge ,Low Pressure Gauge ,Pressure Transmitter,Pressure Switches Differential Pressure Gauges /Indicators - (Pneumatic Pressure)	Using Digital Pressure indicator, Calibrator Make: Polltech Instrument & Test pump and 6.5 Digit DMM Calibrator by Comparison method DKD-R6-01	0 to 20 mbar	0.012mbar



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

35 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
175	MECHANICAL-PRESSURE INDICATING DEVICES	Digital/Analog Pressure Gauges,Indicators,Pressure Transmitter,Pressure Switch,Differential Pressure Gauge/Indicators (Pneumatic Pressure)	Using Digital Pressure Gauge & Pneumatic Test Pump , 6.5 Digit DMM Calibrator By comparison method DKD-R6-01	-0.95 bar to 0	0.0007bar
176	MECHANICAL-PRESSURE INDICATING DEVICES	Digital/Analog Pressure Gauges,Indicators,Pressure Transmitter,Pressure Switch,Differential Pressure Gauge/Indicators - (Pneumatic Pressure)	Using Digital Pressure Gauge & Pneumatic Test Pump , 6.5 Digit DMM Calibrator By Comparison Method DKD-R6-01	0 to 10 bar	0.006bar
177	MECHANICAL-PRESSURE INDICATING DEVICES	Digital/Analog Pressure Gauges,Indicators,Pressure Transmitter,Pressure Switch,Differential Pressure Gauge/Indicators - (Pneumatic Pressure)	Using Digital Pressure Gauge & Pneumatic Test Pump , 6.5 Digit DMM Calibrator By Comparison Method DKD-R6-01	0 to 35 bar	0.007bar





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

36 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
178	MECHANICAL-PRESSURE INDICATING DEVICES	Digital/Analog Pressure Gauges,Pressure Indicators,Pressure Transmitter,Pressure Switches - (Hydraulic Pressure)	Using Digital Pressure Gauge & Pneumatic Test Pump, 6.5 Digit DMM Calibrator By Comparison Method DKD-R6-01	350 bar to 1000 bar	0.61bar
179	MECHANICAL-TORQUE GENERATING DEVICES	Torque Screw Driver/Torque Wrench (Type I and Type II)	By using Torque sensor & display unit as per ISO 6789:2017 standard	0 to 10 Nm	1.57%
180	MECHANICAL-TORQUE GENERATING DEVICES	Torque Screw Driver/Torque Wrench (Type I and Type II)	By using Torque Sensor and Display unit as per ISO 6789:2017 standard	0.1 Nm to 1 Nm	1.28%
181	MECHANICAL-TORQUE GENERATING DEVICES	Torque Wrench (Type I and Type II)	By using Torque sensor & Display unit as per ISO 6789:2017 standard	10 Nm to 100 Nm	1.43%
182	MECHANICAL-TORQUE GENERATING DEVICES	Torque Wrench (Type I and Type II)	By using Torque sensor & Display unit as per ISO 6789:2017 standard	100 Nm to 1000 Nm	1.3%



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

37 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
183	MECHANICAL-VOLUME	Glassware like pipettes ,burettes,measuring cylinder,volumetric flask @27°C	Using Weighing balance of cap. 82 g /220g, readability d=0.01mg/ 0.1mg and distilled water by Gravimetric method based on IS/ISO 4787	>50 ml to 100 ml	0.14ml
184	MECHANICAL-VOLUME	Glassware like pipettes ,burettes,measuring cylinder,volumetric flask @27°C	Using Weighing balance of 1 kg capacity and 1 mg readability and distilled water by Gravimetric method based on IS/ISO 4787	100 ml to 500 ml	0.24ml
185	MECHANICAL-VOLUME	Glassware like pipettes ,burettes,measuring cylinder,volumetric flask@27°C	Using Weighing balance of 220g capacity and 0.01mg readability and distilled water by Gravimetric method based on IS/ISO4787	>10 ml to 50 ml	0.03ml
186	MECHANICAL-VOLUME	Glassware like pipettes ,burettes,measuring cylinder,volumetric flask@27°C	Using Weighing balance of cap. 82 g /220g, readability d=0.01mg/ 0.1mg and distilled water by Gravimetric method based on IS/ISO 4787	1 ml to 10 ml	0.01ml



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

38 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
187	MECHANICAL-VOLUME	Micro-pipette @27°C	Using Weighing balance of 5g Capacity and 0.001mg readability and 82 g capacity and 0.01mg readability and distilled water by Gravimetric method based on ISO 8655 parts 6	1 µl to 100 µl	0.015µl
188	MECHANICAL-VOLUME	Micro-pipette@27°C	Using Weighing balance of cap. 82 g /220g, readability d=0.01mg/ 0.1mg and Weighing balance of Cap. 5.1 g, d= 0.001 mg and distilled water by Gravimetric method based on ISO 8655 parts 6	>100 µl to 10 ml	0.6µl
189	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic Weighing Balance of Class I & Coarser with Readability d=0.1µg & Cap. 2.1 g	Using E1 Class Weights as per OIML R-76-1	0 to 2.1 g	0.004mg





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

39 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
190	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic, Weighing Balances and Comparator of Class I and Coarser with Readability d= 0.001 mg	Using E1 Class Weights as per OIML R-76-1	0 to 100 g	0.015mg
191	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic, Weighing Balance and Compactor of Class I & Coarser with Readability d= 0.001 mg	Using E1 Class Weights ser as per OIML R-76-1	0 to 5 g	0.005mg
192	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic, Weighing Balances and Comparator of Class I and Coarser with Readability d= 0.1 mg	Using E1 Class Weights as per OIML R-76-1	0 to 5000 g	0.7mg
193	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic, Weighing Balances and Comparator of Class II and Coarser with Readability d= 10 mg	Using E2 Class Weights as per OIML R-76-1	0 to 25 kg	16mg
194	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic, Weighing Balances and Compactor of Class I & Coarser with Readability d= 0.001 mg	Using E1 Class Weights as per OIML R-76-1	0 to 20 g	0.01mg



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

40 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
195	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic,Weighing Balances and Compactor of Class I & Coarser with Readability d= 0.01 mg	Using E1 Class Weights as per OIML R-76-1	0 to 200 g	0.03mg
196	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic,Weighing Balances and Compactor of Class I and Coarser with Readability d= 0.01 mg	Using E1 Class Weights as per OIML R-76-1	0 to 1000 g	0.14mg
197	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic,Weighing Balances and Compactor of Class II & Coarser with Readability d= 100 mg	Using E2 Class Weights as per OIML R-76-1	0 to 25 kg	200mg
198	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic,Weighing Balances of Class III & Coarser with Readability d= 1g	Using F1 Class Weight as per OIML R-76-1	0 to 100 kg	6g
199	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic,Weighing Balances of Class IV & Coarser with Readability d= 50g	Using F1 Class Weights as per OIML R-76-1	0 to 400 kg	54g



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

41 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
200	MECHANICAL-WEIGHTS	Weights (E2 Accuracy class & coarser)	Using E1 Accuracy Class reference Weights and Mass Comparator of readability 0.001 mg & Cap. 5.1 g as per OIML R111-1 : 2004	1 g	0.004mg
201	MECHANICAL-WEIGHTS	Weights (E2 Accuracy class & coarser)	Using E1 Accuracy Class reference Weights and Mass Comparator of readability 0.01 mg & Cap. 1.02 kg as per OIML R111-1 : 2004	1 kg	0.21mg
202	MECHANICAL-WEIGHTS	Weights (E2 Accuracy Class & Coarser)	Using E1 Accuracy Class reference Weights and Mass Comparator of readability 0.001 mg & Cap. 5.1 g as per OIML R111-1 : 2004	1 mg	0.002mg
203	MECHANICAL-WEIGHTS	Weights (E2 Accuracy class & coarser)	Using E1 Accuracy Class reference Weights and Mass Comparator of readability 0.01 mg & Cap. 1.02 kg as per OIML R111-1 : 2004	10 g	0.02mg





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

42 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)( $\pm$ )
204	MECHANICAL-WEIGHTS	Weights (E2 Accuracy class & coarser)	Using E1 Class Weights and Mass Comparator of readability 0.001mg- Calibration of Weights class E2 Accuracy & Coarser as per OIML R-111	10 mg	0.002mg
205	MECHANICAL-WEIGHTS	Weights (E2 Accuracy class & coarser)	Using E1 Accuracy Class reference Weights and Mass Comparator of readability 0.01 mg & Cap. 1.02 kg as per OIML R111-1 : 2004	100 g	0.03mg
206	MECHANICAL-WEIGHTS	Weights (E2 Accuracy class & coarser)	Using E1 Accuracy Class reference Weights and Mass Comparator of readability 0.001 mg & Cap. 5.1 g as per OIML R111-1 : 2004	100 mg	0.002mg
207	MECHANICAL-WEIGHTS	Weights (E2 Accuracy class & coarser)	Using E1 Accuracy Class reference Weights and Mass Comparator of readability 0.001 mg & Cap. 5.1 g as per OIML R111-1 : 2004	2 g	0.004mg



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

43 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)( $\pm$ )
208	MECHANICAL-WEIGHTS	Weights (E2 Accuracy class & coarser)	Using E1 Accuracy Class reference Weights and Mass Comparator of readability 0.1 mg & Cap. 5.05 kg as per OIML R111-1 : 2004	2 kg	0.5mg
209	MECHANICAL-WEIGHTS	Weights (E2 Accuracy Class & coarser)	Using E1 Accuracy Class reference Weights and Mass Comparator of readability 0.001 mg & Cap. 5.1 g as per OIML R111-1 : 2004	2 mg	0.002mg
210	MECHANICAL-WEIGHTS	Weights (E2 Accuracy class & coarser)	Using E1 Accuracy Class reference Weights and Mass Comparator of readability 0.01 mg & Cap. 1.02 kg as per OIML R111-1 : 2004	20 g	0.02mg
211	MECHANICAL-WEIGHTS	Weights (E2 Accuracy class & coarser)	Using E1 Accuracy Class reference Weights and Mass Comparator of readability 0.001 mg & Cap. 5.1 g as per OIML R111-1 : 2004	20 mg	0.002mg



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :** ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE, TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-2301 **Page No** 44 of 82

**Validity** 19/05/2022 to 12/10/2023 **Last Amended on** -

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)( $\pm$ )
212	MECHANICAL-WEIGHTS	Weights (E2 Accuracy class & coarser)	Using E1 Accuracy Class reference Weights and Mass Comparator of readability 0.01 mg & Cap. 1.02 kg as per OIML R111-1 : 2004	200 g	0.05mg
213	MECHANICAL-WEIGHTS	Weights (E2 Accuracy class & coarser)	Using E1 Accuracy Class reference Weights and Mass Comparator of readability 0.001 mg & Cap. 5.1 g as per OIML R111-1 : 2004	200 mg	0.002mg
214	MECHANICAL-WEIGHTS	Weights (E2 Accuracy class & coarser)	Using E1 Accuracy Class reference Weights and Mass Comparator of readability 0.001 mg & Cap. 5.1 g as per OIML R111-1 : 2004	5 g	0.005mg
215	MECHANICAL-WEIGHTS	Weights (E2 Accuracy class & coarser)	Using E1 Accuracy Class reference Weights and Mass Comparator of readability 0.1 mg & Cap. 5.05 kg as per OIML R111-1 : 2004	5 kg	0.8mg





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

45 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
216	MECHANICAL-WEIGHTS	Weights (E2 Accuracy class & coarser)	Using E1 Accuracy Class reference Weights and Mass Comparator of readability 0.001 mg & Cap. 5.1 g as per OIML R111-1 : 2004	5 mg	0.002mg
217	MECHANICAL-WEIGHTS	Weights (E2 Accuracy class & coarser)	Using E1 Accuracy Class reference Weights and Mass Comparator of readability 0.01 mg & Cap. 1.02 kg as per OIML R111-1 : 2004	50 g	0.02mg
218	MECHANICAL-WEIGHTS	Weights (E2 Accuracy class & coarser)	Using E1 Accuracy Class reference Weights and Mass Comparator of readability 0.001 mg & Cap. 5.1 g as per OIML R111-1 : 2004	50 mg	0.002mg
219	MECHANICAL-WEIGHTS	Weights (E2 Accuracy class & coarser)	Using E1 Accuracy Class reference Weights and Mass Comparator of readability 0.01 mg & Cap. 1.02 kg as per OIML R111-1 : 2004	500 g	0.1mg



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

46 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
220	MECHANICAL-WEIGHTS	Weights (E2 Accuracy class & coarser)	Using E1 Accuracy Class reference Weights and Mass Comparator of readability 0.001 mg & Cap. 5.1 g as per OIML R111-1 : 2004	500 mg	0.004mg
221	MECHANICAL-WEIGHTS	Weights (F1 Accuracy class & coarser)	Using E2 Accuracy Class reference Weights and Mass Comparator of readability 0.01 g & Cap. 25.5 kg as per OIML R111-1 : 2004	10 kg	10mg
222	MECHANICAL-WEIGHTS	Weights (F1 Accuracy class & coarser)	Using E2 Accuracy Class reference Weights and Mass Comparator of readability 0.01 g & Cap. 25.5 kg as per OIML R111-1 : 2004	20 kg	13mg
223	THERMAL-SPECIFIC HEAT & HUMIDITY	Environmental chamber, Control Storage Room ,Cold Room (Multi position calibration Using 9 Nos. Sensor)	By Using Wireless Data Logger (100 Nos.) at multiposition calibration by spatial mapping comparison method	10 °C to 50 °C	2.30°C



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :** ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE, TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-2301 **Page No** 47 of 82

**Validity** 19/05/2022 to 12/10/2023 **Last Amended on** -

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
224	THERMAL-SPECIFIC HEAT & HUMIDITY	Humidity Calibrator /Generator/Environmental chamber, Control Storage Room (Multi position calibration by Max 100 Nos. Sensor) @25 °C	Using Wireless Data Logger at multiposition calibration by spatial mapping comparison method	10 % to 95 %	4%RH
225	THERMAL-SPECIFIC HEAT & HUMIDITY	Humidity Calibrator /Generator/Environmental chamber, Control Storage Room (Multi position calibration Using upto max. 09 Nos. Sensors ) @25°C	Using multipoint Data logger Yokogawa Mdel:GP 10 and Rotronic Make RH &Temperature Probe sensors & at multiposition calibration by spatial mapping comparison method	10 %RH to 95 %RH	3.9%RH
226	THERMAL-SPECIFIC HEAT & HUMIDITY	Indicator of Humidity calibrator /Generator/Environmental Chamber (Single point calibration) @25°C	Rotronic RH &Temp Probe with Indicator by comparison method .	10 %RH to 90 %RH	0.51% RH





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :** ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE, TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-2301 **Page No** 48 of 82

**Validity** 19/05/2022 to 12/10/2023 **Last Amended on** -

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
227	THERMAL-SPECIFIC HEAT & HUMIDITY	Relative Humidity Sensors with Indicator ,Hygrometer /Humidity Data Logger Humidity Transmitter with sensor @25°C	Using MICHELL RH Generator & Rotronic RH & Temp Probe with Indicator By Comparison Method	10 %RH to 90 %RH	0.54%RH
228	THERMAL-SPECIFIC HEAT & HUMIDITY	Relative Humidity Sensors with Indicator ,Hygrometer /Humidity Data Logger Humidity Transmitter with sensor @50%RH	Using MICHELL RH Generator & Rotronic RH & Temp Prob with Indicator By comparison Method	10 °C to 50 °C	0.35°C
229	THERMAL-TEMPERATURE	Temperature Furnace at Multi Positions calibration by using max. upto 09 sensors)	Using Multipoint Data Logger Yokogawa,Model:GP 10 and 3 wire RTD (PT-100) SENSORS & at multi position calibration by comparison method	25 °C to 400 °C	1.5 °C



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

49 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
230	THERMAL-TEMPERATURE	Environmental Chamber at Multi Positions calibration by using max. upto 09 sensors)	Using Multipoint Data Logger Yokogawa,Model:GP 10 and 3 wire RTD (PT-100) SENSORS & at multi position calibration by comparison method	(-)80 °C to 20 °C	1.5°C
231	THERMAL-TEMPERATURE	Glass Thermometer	Using SPRT with Model:935-14-95L & 934-14-95H WITH Millik Precision Thermometer Temperature Source:Stirred Liquid Bath ,Model:Hydra 798 L By Comparison Method	(-)80 °C to 0	0.08°C
232	THERMAL-TEMPERATURE	Glass Thermometer	Using SPRT with Model:935-14-95L &934-14-95H with Millik Precision Thermometer Temperature Source: Stirred Liquid Bath Model:Hydra 798 By Comparison Method	0 to 300 °C	0.6°C



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :** ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE, TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-2301 **Page No** 50 of 82

**Validity** 19/05/2022 to 12/10/2023 **Last Amended on** -

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
233	THERMAL-TEMPERATURE	Indicator of Liquid Bath, Oil Bath, Dry Block Calibrators Furnace .	Using SPRT with Model:935-14-95L& 34-14-95H with Millik Preci9sion Thermometer (Single Point Calibration) By comparison Method.	(-80) °C to 300 °C	0.06°C
234	THERMAL-TEMPERATURE	Indicator of Oven, Heating Chambers, Water Bath	Using SPRT with Model:935-14-95L& 34-14-95H with Millik Preci9sion Thermometer (Single Point Calibration) By comparison Method	25 °C to 300 °C	0.06 °C
235	THERMAL-TEMPERATURE	Indicator of oven, Muffle Furnace, Heating chambers, Dry Block Calibrators Furnace	Using SPRT with Model:935-14-95L& 34-14-95H with Millik Preci9sion Thermometer (Single Point Calibration) By comparison Method (Single Point Calibration)	300 °C to 600 °C	0.2°C





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

51 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)( $\pm$ )
236	THERMAL-TEMPERATURE	RTD/PRT/ Thermocouple Sensor with or without Indicator ,Temperature Transmitter with Sensor ,Recorder with Sensor	Using SPRT with Model:935-14-95L & 934-14-95H With Millik Precision Thermometer Temperature Source:Stirred Liquid Bath Model:Hydra 798 L By Comparison Method	(-)80 °C to 0	0.016°C
237	THERMAL-TEMPERATURE	RTD/PRT/ Thermocouple Sensor with or without Indicator ,Temperature Transmitter with Sensor ,Recorder with Sensor	Using SPRT with Model:935-14-95L & 934-14-95H With Millik Precision Thermometer Temperature Source:Stirred Liquid Bath Model:Hydra 798 L By Comparison Method	0 to 300 °C	0.02°C
238	THERMAL-TEMPERATURE	RTD/PRT/ Thermocouple Sensor with or without Indicator ,Temperature Transmitter with Sensor ,Recorder with Sensor	Using SPRT with Model 934-14-95H with Milik Precision Thermometer Temperature Source:Dry block calibrator Model:Pegasus 4853 By Comparison Method	300 °C to 600 °C	0.2°C



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

<b>Laboratory Name :</b>	ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE, TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA		
<b>Accreditation Standard</b>	ISO/IEC 17025:2017		
<b>Certificate Number</b>	CC-2301	<b>Page No</b>	52 of 82
<b>Validity</b>	19/05/2022 to 12/10/2023	<b>Last Amended on</b>	-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
239	THERMAL-TEMPERATURE	RTD/PRT/with or without Indicator ,Temperature Transmitter with Sensor ( Fixed Point )	Using SPRT with Model:935-14-95L & 934-14-95H WITH Millik Precision Thermometer Temperature Source: Liquid N2 Bath By Comparison Method	-196 °C	0.02°C
240	THERMAL-TEMPERATURE	Temperature Indicator of Muffle Furnace,Dry Block Calibrator (Using single point calibration)	Using R Type Thermometer with Millik Precision Thermometer By Comparison Method (Single Point Calibration)	1100 °C to 1190 °C	1.6°C
241	THERMAL-TEMPERATURE	Temperature Indicator of Muffle Furnace,Dry Block Calibrator (Using single point calibration)	Using R Type Thermometer with Millik Precision Thermometer By Comparison Method (Single Point Calibration)	600 °C to 1100 °C	1.35°C



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

<b>Laboratory Name :</b>	ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE, TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA		
<b>Accreditation Standard</b>	ISO/IEC 17025:2017		
<b>Certificate Number</b>	CC-2301	<b>Page No</b>	53 of 82
<b>Validity</b>	19/05/2022 to 12/10/2023	<b>Last Amended on</b>	-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
242	THERMAL-TEMPERATURE	Thermocouple with or without indicator Temperature Transmitter with sensor, Thermistor	Using R type Thermometer with Millik Precision Thermometer Temperature Source:Dry Block calibrator,Model:Pegasus 4853 By Comparison Method	1100 °C to 1190 °C	1.5°C
243	THERMAL-TEMPERATURE	Thermocouple with or without indicator Temperature Transmitter with sensor, Thermistor	Using R Type Thermocouple with Millik Precision Thermometer Temperature Source:Dry Block Calibrator ,Model:Pegasus 4853 By Comparison Method	600 °C to 1100 °C	1.35°C





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

54 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
Site Facility					
1	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @ 10 kHz	By using Reference Multimeter Fluke-8508A by Direct Method	1 A to 20 A	0.11 % to 0.30 %
2	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @ 50 Hz to 1 kHz,	By using Reference Multimeter Fluke-8508A by Direct Method	1 A to 20 A	0.10 % to 0.12 %
3	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @ 50 Hz to 10 kHz,	By using Reference Multimeter Fluke-8508A by Direct Method	100 µA to 1 mA	0.08 % to 0.058 %
4	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @ 50 Hz to 10 kHz,	By using Reference Multimeter Fluke-8508A by Direct Method	100 mA to 1 A	0.060 % to 0.11 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

55 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
5	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @ 50 Hz to 10 kHz,	By using Reference Multimeter Fluke-8508A by Direct Method	1 mA to 100 mA	0.058 % to 0.060 %
6	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @10 Hz,	By using Reference Multimeter Fluke-8508A by Direct Method	100 mA to 1 A	0.06 % to 0.11 %
7	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @10 Hz,	By using Reference Multimeter Fluke-8508A by Direct Method	100 $\mu$ A to 100 mA	0.08 % to 0.06 %
8	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Power & Active Energy Single Phase & Three Phase @ 50 Hz, 240 V/ 1 mA to 100 A at UPF, 0.5PF & 0.8 PF	By using Reference Energy Meter ZERA-MT310 by Comparison Method	0.12 W to 72 kW	0.076%
9	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Resistance @ 1 kHz	By using RLC Digibridge GenRad 1658 by Direct Method	100 Ohm to 10 k Ohm	0.12 % to 0.42 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

56 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
10	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Resistance @1 kHz	By using RLC Digibridge GenRad 1658 by Direct Method	1 Ohm to 100 Ohm	0.12%
11	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 10 Hz	By using Reference Multimeter Fluke-8508A by Direct Method	10 mV to 10 V	0.08 % to 0.036 %
12	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 10 Hz	By using Reference Multimeter Fluke-8508A by Direct Method	10 V to 100 V	0.036 % to 0.044 %
13	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 10 Hz,	By using Reference Multimeter Fluke-8508A by Direct Method	1 mV to 10 mV	0.93 % to 0.08 %
14	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 10 kHz	By using Reference Multimeter Fluke-8508A by Direct Method	1 mV to 10 mV	0.62 % to 0.069 %





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

57 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
15	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 10 kHz,	By using Reference Multimeter Fluke-8508A by Direct Method	10 mV to 100 mV	0.069 % to 0.017 %
16	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 10 kHz,	By using Reference Multimeter Fluke-8508A by Direct Method	100 mV to 100 V	0.017 % to 0.013 %
17	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 100 kHz,	By using Reference Multimeter Fluke-8508A by Direct Method	10 mV to 100 mV	0.33 % to 0.12 %
18	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 100 kHz,	By using Reference Multimeter Fluke-8508A by Direct Method	100 mV to 10 V	0.12 % to 0.031 %
19	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 50 Hz to 1 kHz,	By using Reference Multimeter Fluke-8508A by Direct Method	1 mV to 100 mV	0.57 % to 0.02 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

58 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
20	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 50 Hz to 1 kHz,	By using Reference Multimeter Fluke-8508A by Direct Method	100 mV to 100 V	0.020 % to 0.013 %
21	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @ 50 Hz to 10 kHz	By using Reference Multimeter Fluke-8508A by Direct Method	100 V to 1000 V	0.013 % to 0.017 %
22	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	Capacitance @ 1 kHz	By using Decade Capacitance Box Vaiseshikha-Type 7500 by Direct Method	10 nF	1.16 %
23	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	Capacitance@1 kHz	By using RLC Digibridge GenRad 1658 by Direct/ Comparison Method	1 nF to 100 $\mu$ F	0.12 %
24	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	Inductance @1 kHz	By using RLC Digibridge GenRad 1658 by Direct / Comparison Method	1 mH to 10 H	0.19 % to 0.13 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

59 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
25	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	Inductance @1 kHz	By Using RLC Digibridge GenRad 1658 by Direct Method	100 $\mu$ H to 1 mH	0.32 % to 0.19 %
26	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Current @ 50 Hz to 1 kHz	Using Multifunction Calibrator,Fluke-550 2 A By Direct Method	30 $\mu$ A to 100 $\mu$ A	0.56 % to 0.29 %
27	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Current @ 50 Hz to 1 kHz	By using MFC 5502A by Direct / comparison Method	1 A to 10 A	0.08 % to 0.14 %
28	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Current @ 50 Hz to 1 kHz	By using MFC 5502A by Direct/ comparison method	10 A to 20 A	0.14 % to 0.2 %
29	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Current @ 50 Hz to 1 kHz	By using MFC 5502A by Direct / comparison method	100 mA to 1 A	0.075 % to 0.05 %





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

60 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
30	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Current @ 50 Hz to 1 kHz	By using MFC 5502A Direct method	100 mA to 1 A	0.075 % to 0.08 %
31	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Current @ 50 Hz,	Using Multifunction Calibrator,Fluke-550 2 A & Current Coil, Fluke By Direct Method.	20 A to 1000 A	0.82 % to 0.85 %
32	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Current @ 50Hz to 1 kHz,	Using Multifunction Calibrator,Fluke-550 2 A By Direct Method.	1 mA to 100 mA	0.14 % to 0.08 %
33	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Current @ 50Hz to 1kHz	Using Multifunction Calibrator,Fluke-550 2 A By Direct Method.	100 µA to 1 mA	0.29 % to 0.14 %
34	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Power (1 Phase) @UPF, 50 Hz, 30V to 320V/ 1 mA to 20A	Using Multifunction Multiproduct Calibrator ,Fluke-5502 A +50 Turn By Direct Method.	30 mW to 6.4 kW	0.24 % to 0.07 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

61 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)( $\pm$ )
35	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Voltage @ 50Hz to 1kHz,	Using Multifunction Calibrator, Fluke -5502A By Direct Method.	10 mV to 100 mV	0.36 % to 0.09 %
36	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Voltage @ 50Hz to 1kHz,	Using Multifunction Calibrator ,Fluke -5502A By Direct / Comparison Method.	10 V to 1000 V	0.060 % to 0.064 %
37	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Voltage @ 50Hz to 1kHz,	Using Multifunction Calibrator ,Fluke -5502A By Direct Method.	100 mV to 10 V	0.09 % to 0.06 %
38	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	Capacitance @ 1 kHz	By using Decade Capacitance Box Vaiseshikha-Type 7500 by Direct Method	1 nF	1.16 %
39	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	Capacitance @ 1 kHz	By using Decade Capacitance Box Vaiseshikha-Type 7500 by Direct Method	100 nF	1.16 %
40	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	Capacitance @ 1 kHz,	By using Decade Capacitance Box Vaiseshikha-Type 7500 by Direct / Comparison Method	1 $\mu$ F	1.16 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

62 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
41	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	Capacitance @ 100 Hz	By using Decade Capacitance Box Vaiseshikha-Type 7500 by Direct Method	1 $\mu$ F	1.16 %
42	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	Capacitance @ 100 Hz	By using Decade Capacitance Box Vaiseshikha-Type 7500 by Direct Method	1 mF	1.17 %
43	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	Capacitance @ 100 Hz	By using Decade Capacitance Box Vaiseshikha-Type 7500 by Direct Method	10 $\mu$ F	1.16 %
44	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	Capacitance @ 100 Hz	By using Decade Capacitance Box Vaiseshikha-Type 7500 by Direct Method	100 $\mu$ F	1.17%
45	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	Capacitance@ 1kHz	By using MFC Fluke 5502A by Direct Method	1 nF to 100 $\mu$ F	2.0 % to 1.91 %
46	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	Inductance @1 kHz	using Decade inductance box Vaiseshika/7500 by Direct method	1 H	1.16 %





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

63 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
47	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	Inductance @1 kHz	By using Decade inductance box Vaiseshika/7500 by Direct method	1 mH	1.16 %
48	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	Inductance @1 kHz	using Decade inductance box Vaiseshika/7500 by Direct/ Comparison method	10 H	1.16 %
49	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	Inductance @1 kHz	using Decade inductance box Vaiseshika/7500 by Direct/ Comparison method	10 mH	1.16 %
50	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	Inductance @1 kHz	By using Decade inductance Box Vaiseshikha/ 7500 by Direct Method	100 µH	1.16 %
51	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	Inductance @1 kHz	using Decade inductance box Vaiseshika/7500 by Direct method	100 mH	1.16 %
52	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	DC Current	By using Reference Multimeter Fluke-8508A by Direct Method	1 A to 10 A	0.023 % to 0.09 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

64 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
53	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Current	By using Reference Multimeter Fluke-8508A by Direct Method	10 $\mu$ A to 10 mA	0.009 % to 0.0029 %
54	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Current	By using Reference Multimeter Fluke-8508A by Direct Method	10 A to 20 A	0.051 % to 0.049 %
55	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Current	By using Reference Multimeter Fluke-8508A by Direct Method	10 mA to 100 mA	0.0029 % to 0.007 %
56	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Current	By using Reference Multimeter Fluke-8508A by Direct Method	100 mA to 1 A	0.007 % to 0.023 %
57	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	Using Reference Multimeter Fluke 8508A By Direct Method	0.1 mV to 1 mV	0.13 % to 0.016 %
58	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	Using Multifunction Fluke 80K-40 HV Probe with DMM 87 V by Direct method	1 kV to 40 kV	1.84%



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

65 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
59	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	By using Reference Multimeter Fluke 8508A by Direct Method	1 mV to 10 mV	0.016 % to 0.004 %
60	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	By using Reference Multimeter Fluke 8508 A by Direct Method	1 V to 1000 V	0.0007 % to 0.001 %
61	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	using Reference Multimeter Fluke -8508A By Direct Method	10 mV to 100 mV	0.004 % to 0.001 %
62	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	By using Reference Multimeter Fluke-8508A by Direct Method	100 mV to 1 V	0.001 % to 0.0007 %
63	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance	By using Reference Multimeter Fluke-8508A by Direct Method	1 M ohm to 10 M ohm	0.002 % to 0.0043 %
64	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance	By using Reference Multimeter Fluke-8508A by Direct / Comparison Method	1 m ohm to 10 m ohm	0.47 % to 0.049 %





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

66 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
65	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance	By using Reference Multimeter Fluke-8508A by Direct Method	1 ohm to 10 ohm	0.0063 % to 0.0017 %
66	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance	By using Reference Multimeter Fluke-8508A by Direct Method	10 k ohm to 1 M ohm	0.001 % to 0.002 %
67	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance	By using Reference Multimeter Fluke-8508A by Direct Method	10 m ohm to 100 m ohm	0.052 % to 0.0071 %
68	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance	By using Reference Multimeter Fluke-8508A by Direct Method	10 M ohm to 100 M ohm	0.0043 % to 0.026 %
69	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance	By using Reference multimeter Fluke 8508A by Direct Method	10 Ohm to 10 k Ohm	0.0017 % to 0.001 %
70	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance	By using Reference Multimeter Fluke-8508A by Direct Method	100 m ohm to 1 ohm	0.0071 % to 0.0063 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

67 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)( $\pm$ )
71	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance	By using Reference Multimeter Fluke-8508A by Direct Method	100 M ohm to 10 G ohm	0.026 % to 0.29 %
72	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using Multifunction Calibrator - Fluke-5502A By Direct Method.	1 mA to 100 mA	0.02 % to 0.017 %
73	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Resistance	By using MFC Fluke 5502A by Direct Method	1 M Ohm to 10 M Ohm	0.019 % to 0.07 %
74	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Resistance	By using MFC Fluke 5502A by Direct Method	100 m Ohm to 1 Ohm	0.8 % to 0.58 %
75	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	High Resistance	By using Decade Resistor Box Vaiseshikha-Type 8400 HVSD(Standard) by Direct Method	1 G ohm	1.41 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

68 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
76	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	High Resistance	By using Decade Resistor Box Vaiseshikha-Type 8400 HVSD(Standard) by Direct Method	1 k ohm	0.6 %
77	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	High Resistance	By using Decade Resistor Box Vaiseshikha-Type 8400 HVSD(Standard) by Direct Method	1 M Ohm	1.0 %
78	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	High Resistance	By using Decade Resistor Box Vaiseshikha-Type 8400 HVSD(Standard) by Direct Method	1 T ohm	8.33 %
79	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	High Resistance	By using Decade Resistor Box Vaiseshikha-Type 8400 HVSD(Standard) by Direct Method	10 G ohm	1.7 %
80	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	High Resistance	By using Decade Resistor Box Vaiseshikha-Type 8400 HVSD(Standard) by Direct Method	10 k ohm	0.06 %





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

69 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
81	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	High Resistance	By using Decade Resistor Box Vaiseshikha-Type 8400 HVSD(Standard) by Direct Method	10 M OHM	1.4 %
82	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	High Resistance	By using Decade Resistor Box Vaiseshikha-Type 8400 HVSD(Standard) by Direct Method	100 G ohm	3.8 %
83	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	High Resistance	By using Decade Resistor Box Vaiseshikha-Type 8400 HVSD(Standard) by Direct Method	100 k ohm	0.6 %
84	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	High Resistance	By using Decade Resistor Box Vaiseshikha-Type 8400 HVSD(Standard) by Direct Method	100 M ohm	1.4 %
85	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Low Resistance	By using Micro/Milli Ohm Meter Calibrator Vaiseshikha-9409 CAL by Direct Method	1 m ohm	0.16 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

70 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
86	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Low Resistance	By using Micro/Milli Ohm Meter Calibrator Vaiseshikha-9409 CAL by Direct Method	1 ohm	0.07 %
87	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Low Resistance	By using Micro/Miliohm meter calibrator Vaiseshika -9409 Cal by Direct Method	10 $\mu$ Ohm	3.2 %
88	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Low Resistance	By using Micro/Milli Ohm Meter Calibrator Vaiseshikha-9409 CAL by Direct Method	10 $\mu$ ohm	3.48 %
89	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Low Resistance	By using Micro/Milli Ohm Meter Calibrator Vaiseshikha-9409 CAL by Direct Method	10 m ohm	0.08 %
90	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Low Resistance	By using Micro/Milli Ohm Meter Calibrator Vaiseshikha-9409 CAL by Direct/Comparison Method	100 $\mu$ ohm to	1.6 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

71 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
91	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Low Resistance	By using Micro/Milli Ohm Meter Calibrator Vaiseshikha-9409 CAL by Direct Method	100 m ohm	0.07 %
92	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Low Resistance	By using Micro/Milli Ohm Meter Calibrator Vaiseshikha-9409 CAL by Direct Method	2 ohm	0.07 %
93	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance	By using Decade Resistor Box Vaiseshikha-Type 8400 HVSD(Standard) by Direct Method	100 M Ohm	1.4 %
94	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Digital/Analogue stop watch, Digital Timer	Using Time Interval Meter by Comparison Method	1 hr to 5 hr	6.65 s to 24.62 s
95	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Digital/Analogue stop watch, Digital Timer	Using Time Interval Meter by Comparison Method	1 s to 60 s	0.14 s to 0.16 s





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

72 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
96	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Digital/Analogue Stop watch, Digital Timer	Using Time Interval Meter by Comparison Method	200 ms to 1 s	0.07 s to 0.14 s
97	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Digital/Analogue stop watch, Digital Timer	Using Time Interval Meter by Comparison Method	60 s to 1 hr	0.16 s to 6.65 s
98	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Frequency measurement for MFC's, DMM	By using Reference Multimeter Fluke-8508A by Direct / Comparison Method	10 Hz to 1 MHz	0.0013 % to 0.0012 %
99	ELECTRO-TECHNICAL-TIME & FREQUENCY (Source)	Frequency meter, DMM	Using Multifunction Multiproduct Calibrator, Fluke 5502A By Direct Method.	10 Hz to 1 MHz	0.062 % to 0.0021 %
100	MECHANICAL-ACCELERATION AND SPEED	RPM Indicator	Using Digital Tachometer & Mechanical type Tachometer Calibrator By Comparison Method	10 rpm to 20000 rpm	2.0rpm



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

73 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
101	MECHANICAL-ACCELERATION AND SPEED	Tachometer (contact Mode)	Using Digital Tachometer & Mechanical type Tachometer Calibrator By Comparison Method	10 rpm to 1000 rpm	3.4rpm
102	MECHANICAL-ACCELERATION AND SPEED	Tachometer (contact Mode)	Using Digital Tachometer & Mechanical type Tachometer Calibrator By Comparison Method	1000 rpm to 5000 rpm	6.6rpm
103	MECHANICAL-ACCELERATION AND SPEED	Tachometer (Noncontact Type)	Using Digital Tachometer & Mechanical type Tachometer Calibrator By Comparison Method	10 rpm to 100 rpm	3rpm
104	MECHANICAL-PRESSURE INDICATING DEVICES	Digital Pressure Indicator ,Manometer Magnahelic Gauge ,Low Pressure Gauge ,Pressure Transmitter,Pressure Switches Differential Pressure Gauges /Indicators - (Pneumatic Pressure)	Using Digital Pressure indicator, Calibrator Make: Polltech Instrument & Test pump , 6.5 Digit DMM Calibrator by Comparison method DKD-R6-01	0 to 200 mbar	0.1mbar



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

74 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)( $\pm$ )
105	MECHANICAL-PRESSURE INDICATING DEVICES	Digital/Analog Pressure Gauges,Indicators,Pressure Transmitter,Pressure Switch,Differential Pressure Gauge/Indicators - (Pneumatic Pressure)	Using Digital Pressure Gauge & Pneumatic Test Pump, 6.5 Digit DMM Calibrator by Comparison Method DKD-R6-01	0 to 2 bar	0.0006bar
106	MECHANICAL-PRESSURE INDICATING DEVICES	Digital Pressure Indicator ,Manometer Magnahelic Gauge ,Low Pressure Gauge ,Pressure Transmitter,Pressure Switches Differential Pressure Gauges /Indicators - (Pneumatic Pressure)	Using Digital Pressure indicator, Calibrator Make: Polltech Instrument & Test pump and 6.5 Digit DMM Calibrator by Comparison method DKD-R6-01	0 to 20 mbar	0.012mbar
107	MECHANICAL-PRESSURE INDICATING DEVICES	Digital/Analog Pressure Gauges,Indicators,Pressure Transmitter,Pressure Switch,Differential Pressure Gauge/Indicators - (Pneumatic Pressure)	Using Digital Pressure Gauge & Pneumatic Test Pump , 6.5 Digit DMM Calibrator By Comparison Method DKD-R6-01	0 to 10 bar	0.006bar





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

75 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)( $\pm$ )
108	MECHANICAL-PRESSURE INDICATING DEVICES	Digital/Analog Pressure Gauges,Indicators,Pressure Transmitter,Pressure Switch,Differential Pressure Gauge/Indicators - (Pneumatic Pressure)	Using Digital Pressure Gauge & Pneumatic Test Pump , 6.5 Digit DMM Calibrator By Comparison Method DKD-R6-01	0 to 35 bar	0.007bar
109	MECHANICAL-PRESSURE INDICATING DEVICES	Digital/Analog Pressure Gauges,Pressure Indicators,Pressure Transmitter,Pressure Switches - (Hydraulic Pressure)	Using Digital Pressure Gauge & Pneumatic Test Pump, 6.5 Digit DMM Calibrator By Comparison Method DKD-R6-01	350 bar to 1000 bar	0.61bar
110	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic,Weighing Balance and Compator of Class I & Coarser with Readability d= 0.001 mg	Using E1 Class Weights ser as per OIML R-76-1	0 to 5 g	0.005mg
111	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic,Weighing Balances and Comparator of Class I and Coarser with Readability d= 0.1 mg	Using E1 Class Weights as per OIML R-76-1	0 to 5000 g	0.7mg



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :** ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE, TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-2301 **Page No** 76 of 82

**Validity** 19/05/2022 to 12/10/2023 **Last Amended on** -

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
112	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic,Weighing Balances and Comparator of Class II and Coarser with Readability d= 10 mg	Using E2 Class Weights as per OIML R-76-1	0 to 25 kg	16mg
113	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic,Weighing Balances and Compator of Class I & Coarser with Readability d= 0.001 mg	Using E1 Class Weights as per OIML R-76-1	0 to 20 g	0.01mg
114	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic,Weighing Balances and Compator of Class I & Coarser with Readability d= 0.01 mg	Using E1 Class Weights as per OIML R-76-1	0 to 200 g	0.03mg
115	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic,Weighing Balances and Compator of Class II & Coarser with Readability d= 100 mg	Using E2 Class Weights as per OIML R-76-1	0 to 25 kg	200mg
116	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic,Weighing Balances of Class III & Coarser with Readability d= 1g	Using F1 Class Weight as per OIML R-76-1	0 to 100 kg	6g



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

77 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)( $\pm$ )
117	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic,Weighing Balances of Class IV & Coarser with Readability d= 50g	Using F1 Class Weights as per OIML R-76-1	0 to 400 kg	54g
118	THERMAL-SPECIFIC HEAT & HUMIDITY	Environmental chamber, Control Storage Room ,Cold Room (Multi position calibration Using 9 Nos. Sensor)	By Using Wireless Data Logger (100 Nos.) at multiposition calibration by spatial mapping comparison method	10 °C to 50 °C	2.30°C
119	THERMAL-SPECIFIC HEAT & HUMIDITY	Humidity Calibrator /Generator/Environm ental chamber, Control Storage Room (Multi position calibration by Max 100 Nos. Sensor) @25 °C	Using Wireless Data Logger at multiposition calibration by spatial mapping comparison method	10 % to 95 %	4%RH
120	THERMAL-SPECIFIC HEAT & HUMIDITY	Humidity Calibrator /Generator/Environm ental chamber, Control Storage Room (Multi position calibration Using upto max. 09 Nos. Sensors ) @25°C	Using multipoint Data logger Yokogawa Mdel:GP 10 and Rotronic Make RH &Temperature Probe sensors & at multiposition calibration by spatial mapping comparison method	10 %RH to 95 %RH	3.9%RH





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

78 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
121	THERMAL-SPECIFIC HEAT & HUMIDITY	Relative Humidity Sensors with Indicator ,Hygrometer /Humidity Data Logger Humidity Transmitter with sensor @25°C	Using MICHELL RH Generator & Rotronic RH & Temp Probe with Indicator By Comparison Method	10 %RH to 90 %RH	0.54%RH
122	THERMAL-TEMPERATURE	Temperature Furnace at Multi Positions calibration by using max. upto 09 sensors)	Using Multipoint Data Logger Yokogawa,Model:GP 10 and 3 wire RTD (PT-100) SENSORS & at multi position calibration by comparison method	25 °C to 400 °C	1.5 °C
123	THERMAL-TEMPERATURE	Autoclaves (For Non Medical Purpose Only) at Multi-positions calibration by using Max upto 16 nos sensors.	Using Multipoint Data Logger Yokogawa,Model:GP 10 and 3 wire RTD (PT-100) Sensors by using comparison method	120 °C to 150 °C	1.5°C
124	THERMAL-TEMPERATURE	Environmental Chamber at Multi Positions calibration by using max. upto 09 sensors)	Using Multipoint Data Logger Yokogawa,Model:GP 10 and 3 wire RTD (PT-100) SENSORS & at multi position calibration by comparison method	(- )80 °C to 20 °C	1.5°C



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

79 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
125	THERMAL-TEMPERATURE	Incubator (For Non Medical Purpose Only), BOD Incubator (For Non Medical Purpose Only) at Multi-positions by using Max upto 16 nos sensors.	Using Multipoint Data Logger Yokogawa,Model:GP 10 and 3 wire RTD (PT-100) SENSORS by comparison Method.	5 °C to 60 °C	1.5°C
126	THERMAL-TEMPERATURE	Indicator of Liquid Bath,Oil Bath, Dry Block Calibrators Furnace .	Using SPRT with Model:935-14-95L& 34-14-95H with Millik Preci9sion Thermometer (Single Point Calibration) By comparison Method.	(-80) °C to 300 °C	0.06°C
127	THERMAL-TEMPERATURE	Indicator of Oven, Heating Chambers,Water Bath	Using SPRT with Model:935-14-95L& 34-14-95H with Millik Preci9sion Thermometer (Single Point Calibration) By comparison Method	25 °C to 300 °C	0.06 °C



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

80 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
128	THERMAL-TEMPERATURE	Indicator of oven, Muffle Furnace, Heating chambers, Dry Block Calibrators Furnace	Using SPRT with Model:935-14-95L& 34-14-95H with Millik Preci9sion Thermometer (Single Point Calibration) By comparison Method (Single Point Calibration)	300 °C to 600 °C	0.2°C
129	THERMAL-TEMPERATURE	Refrigerator, Freezers, Deep Freezers at Multi-positions calibration by using Max upto 16 nos sensors.	Using Multipoint Data Logger Yokogawa, Model:GP 10 and 3 wire RTD (PT-100) SENSORS by comparison Method	(-80) °C to 20 °C	1.5°C
130	THERMAL-TEMPERATURE	RTD/PRT/ Thermocouple Sensor with or without Indicator ,Temperature Transmitter with Sensor ,Recorder with Sensor	Using SPRT with Model 934-14-95H with Milik Precision Thermometer Temperature Source:Dry block calibrator Model:Pegasus 4853 By Comparison Method	300 °C to 600 °C	0.2°C





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :** ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE, TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-2301

**Validity** 19/05/2022 to 12/10/2023

**Page No** 81 of 82

**Last Amended on** -

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
131	THERMAL-TEMPERATURE	Temperature Indicator of Muffle Furnace,Dry Block Calibrator (Using single point calibration)	Using R Type Thermometer with Millik Precision Thermometer By Comparison Method (Single Point Calibration)	1100 °C to 1190 °C	1.6°C
132	THERMAL-TEMPERATURE	Temperature Indicator of Muffle Furnace,Dry Block Calibrator (Using single point calibration)	Using R Type Thermometer with Millik Precision Thermometer By Comparison Method (Single Point Calibration)	600 °C to 1100 °C	1.35°C
133	THERMAL-TEMPERATURE	Thermocouple with or without indicator Temperature Transmitter with sensor, Thermistor	Using R type Thermometer with Millik Precision Thermometer Temperature Source:Dry Block calibrator,Model:Pegasus 4853 By Comparison Method	1100 °C to 1190 °C	1.5°C



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

ARCHERCAL PRIVATE LIMITED, B-103, TANVI'S TANISHKA INDUSTRIAL ESTATE,  
TANVI COMPLEX, S. V. ROAD, DAHISAR EAST, MUMBAI, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2301

**Page No**

82 of 82

**Validity**

19/05/2022 to 12/10/2023

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)( $\pm$ )
134	THERMAL-TEMPERATURE	Thermocouple with or without indicator Temperature Transmitter with sensor, Thermistor	Using R Type Thermocouple with Millik Precision Thermometer Temperature Source:Dry Block Calibrator ,Model:Pegasus 4853 By Comparison Method	600 °C to 1100 °C	1.35°C

\* CMCs represent expanded uncertainties expressed at approximately the 95% level of confidence, using a coverage factor of  $k = 2$ .